



## **Simultaneous Determination of Hydrochlorothiazide and Telmisartan by Using Reverse Phase HPLC Technique**

**P. Haritha<sup>1</sup>, B. Sreenivasa Rao\*<sup>2</sup>, B. Naganjaneyulu<sup>2</sup> and Y. Sunandamma<sup>3</sup>**

1. Research Scholar, Department of Chemistry, JNTUK, Kakinada, AP, **INDIA**

2. Department of Chemistry, GITAM University, Visakhapatnam, AP, **INDIA**

3. Department of Chemistry, Vikrama Simhapuri University, Nellore, AP, **INDIA**

Email: [battula\\_sr@gitam.edu](mailto:battula_sr@gitam.edu)

Accepted on 24<sup>th</sup> December 2013

---

### **ABSTRACT**

*A rapid specific reverse-phase HPLC method has been developed for assaying Hydrochlorothiazide and Telmisartan by simultaneous method in pharmaceutical dosage forms. The method involves an isocratic elution of drug in a stationary phase of Phenomenex Prodigy, C18, 150 mm X 4.6 mm, 5  $\mu$ m column using a mobile phase composition of methanol and 0.1 % orthophosphoric acid in the composition ratio 70:30 % (V/V) and with flow rate of 1.0 ml / min at 260nm of detection. The developed method is found to be linear in the range of 4.99 to 99.80  $\mu$ g/ml for Telmisartan and 2.49 to 49.75  $\mu$ g/ml for Hydrochlorothiazide respectively. The injection volume is 20 $\mu$ L. The method has been validated for specificity, linearity, range, precision, accuracy, limit of detection, limit of quantification, ruggedness and robustness. The % recovery of Telmisartan and Hydrochlorothiazide was found to be in the range of 99.00 % - 101.00 %. All the validation parameters are within the acceptance range.*

**Keywords:** Reverse-phase HPLC, Isocratic, Hydrochlorothiazide, Telmisartan.

---